

User Manual

eBlot™ L2

Fast Protein Transfer Device

For High Quality Wet Protein Transfer



Version: V1

Cat. No: L00981

CONTENTS

Important Notice	1
Quick Guide	2
1. Product Overview	5
1.1 Instrument Overview and Features	5
1.2 Components, Specifications and Installation	6
2. Device and Interface Display	9
2.1 Device and Transfer Cassette	9
2.2 Interface Display and Method Setting	10
3. Device Operation and Application Data	15
3.1 Reagent Preparation	15
3.2 Method Selection Guide	16
3.3 Operation Instructions	17
3.4 Application Data	22
4. Device Maintenance	23
5. Troubleshooting	24
6. Consumables Offering	28
7. Related Products	29
8. Warranty	32

Important Notice

Users must read this manual in its entirety before operating eBlot™ L2 Fast Transfer System, while paying close attention to the safety precautions listed in this manual.

Version: V1; Date: 10152024

Please be aware that GenScript will not be held responsible for any personal injury or damage to the instrument caused by failure to follow the instructions provided in the manual.

Safety

Please follow the operation instructions provided in this manual

Do not place the instrument at a location prone to vibrations

eBlot™ L2 is intended for laboratory use only

Place the equipment on a level laboratory bench and ensure clean and ventilated surroundings

Improper operation may result in damage to the equipment

Users should not dismantle the equipment themselves to avoid equipment damage or personal injury

Original packaging must be used for equipment transportation



Warning sign



Electrical shock warning sign

For research use only. Not intended for diagnostic use.

Quick Guide

Instructions for transferring mini or midi gels using eBlot™ L2

* Please read and follow the instructions carefully if you are first-time user of eBlot™ L2

1

Preparation

Reagents

- Diluted eBlot L2 Transfer Buffer (300 mL/Standard method)
- Diluted eBlot L2 Equilibrium Buffer (60 mL per membrane in the plastic container provided)
- Distilled water (250 mL per transfer in the silver tray provided)

Note: Regularly empty the waste container

Other Materials

- Pre-run PAGE gel
- NC membrane or Pre-activated PVDF membrane
- Dry transfer sponges (2 pieces per transfer)


2

Opening of Transfer Cassette

Place the transfer cassette on the bench as shown below.

- a. Release the cassette latch by rotating it upward.
- b. Lift the latch and fully open the cassette cover to approximately > 90°.

Note: The anode side is marked with a red sealing strip.



3


Preparation of Gel and Membrane

Rinse the gel cassette under running water (5 seconds) to remove remaining running buffer. Open the gel cassette and cut off the gel well fingers. Place the gel and membrane in the eBlot L2 Equilibrium buffer for 1 minute*.

Note:

- Cut off any uneven part of the pre-cast gel, or the stacking gel of the home-made gel.
- Pre-wet the PVDF membrane with alcohol before equilibrium.


*The gel equilibrium time should NOT exceed 5 minutes.



4

Assembly of Transfer Sandwich

On the red anode side, place the materials in the following order: dry sponge, equilibrated membrane, equilibrated gel. Use the roller to roll out bubbles between the membrane and the gel, then place another piece of dry sponge on top of the gel.




a. Assembly of 2 mini gels


b. Assembly of 1 midi gel

Note:

- Note the correct size of the components: Sponge $\geq$ Membrane $\geq$ Gel.
- Avoid overlapping or touching between two mini gels.
- Place the gel and membrane close to the bottom without exceeding green indication line that represents the edge of the transfer surface.
- Finish the assembly quickly to prevent from drying out.
- When transfer only 1 mini gel in the midi cassette, place the mini gel in the middle of the same indicated area.



Transfer 2 mini gels in one cassette



Transfer 1 midi gel in one cassette


5

Start Transfer

Close the transfer cassette, and insert the assembled cassette into the selected channel with the scale lines completely covered. Press  to initiate the transfer.

Note:

- The latch and scale lines are facing toward the user.
- Start transfer no longer than 5 minutes after the assembly of transfer sandwich.
- Refer to section 2.2.2 Transfer Method Overview for setting up the methods.



Start the transfer


6

Complete Transfer


The device will beep when the countdown reaches 0. Press  to return to the initial interface. Remove the transfer cassette from the instrument. Open the transfer cassette, then immediately place the transfer sandwich into a tray containing distilled water. Disassemble the transfer sandwich and take out the membrane and continue with the next step as per your normal protocol. After each transfer, rinse the cassette under running water for 1 minute and let it dry.

Note:

If the PVDF membrane is dry after transfer, it must be reactivated in applicable alcohol before proceeding to the subsequent experiments.



Transfer completed



Disassembly of the transfer sandwich

1. Product Overview

1.1 Instrument overview

eBlot™ L2 Fast Protein Transfer Device is the second-generation protein transfer instrument engineered to efficiently handle the transfer of proteins. eBlot™ L2 can simultaneously complete the transfer process for up to 4 mini gels or 2 midi gels in 5-15 minutes. The unique eBlot technology seamlessly combines the benefits of traditional wet transfer and semidry/dry transfer methods, ensuring with very short transfer times producing exceptional transfer efficiency and consistent results for proteins of all sizes.

eBlot™ L2 system consists of one device and two transfer cassettes, each capable of running independently. The transfer process is fully automated using alcohol-free proprietary reagents, ensuring a clean, safe and convenient transfer process. eBlot™ L2 is compatible with commonly used gels, such as Bis-Tris and Tris-Gly gels, supporting versatile research needs.

Features

- Fast wet transfer in 5-15 minutes
- High throughput transfer for up to 4 mini gels or 2 midi gels
- Alcohol-free reagents with ddH₂O dilution
- Result consistency and reproducibility
- Easy operation with touchscreen and fast start
- Versatile compatibilities
 - All size proteins
 - Bis-Tris and Tris-Gly gels
 - PVDF, nitrocellulose membranes

1.2 Components, specifications and installation

Components

eBlot™ L2 Fast Protein Transfer Device includes the following components. Please carefully check the components before installation.

Products	Catalog Number	Quantity
eBlot™ L2 Fast Transfer Device	L00980	1
eBlot L2 Transfer Cassette, midi	L00982	2
Power cord	-	1
Forceps	-	1
Spatula	-	1
Silver tray	-	1
Roller	-	1
Silicon tubing	-	4
Plastic container	-	1
Plastic container L	-	1
Stylus pen	-	1
eBlot L2 reagent bottle cap	-	2
Liquid Container, 5L	-	2

Specifications

eBlot™ L2 Fast Transfer Device

Catalog Number:	L00980
Weight:	7.5 kg
Dimension:	280 mm (L) × 260 mm(W) × 320 mm (H)
Electrical Requirements:	100-240 V, 50/60 Hz, 800 W
Digital Module:	LCD display, alarm and LED light
Application:	Fast transfer of proteins from polyacrylamide gel to PVDF or nitrocellulose membrane
Materials:	ABS, PP, Stainless steel, Plasticized silicone
Operating Temperature:	15-40 °C
Forceps:	Stainless steel
Spatula:	Polycarbonate
Silver tray:	Stainless steel

Note: Avoid contact with acid, alkaline, acetone or any other reagents that might erode or damage the exterior surface of the device.


eBlot L2 Transfer Cassette, midi

Catalog Number:	L00982
Dimension:	205 mm (L) × 160 mm (W) × 40 mm (H)
Weight:	1.9 kg
Material:	Aluminum-magnesium alloy, titanium, stainless steel
Application:	Each cassette can transfer up to 2 mini gels or 1 midi gel

Note: Avoid contact with acid, alkaline, acetone or any other reagents that might erode or damage the exterior surface of cassettes

Installation

Users must follow the instructions below when installing the instrument

Steps	Descriptions
1	Check if the power cord matches with the outlet
2	Place eBlot™ L2 on a levelled laboratory bench
3	Maintain cleanliness and ventilation around the instrument, especially behind the device
4	Ensure the power switch is off
5	Connect the inlet and outlet tubes to the color-matched reagent containers as followed: · Transparent cap - Transfer buffer · Yellow cap - Distilled water · Black cap - Waste
	
6	Connect the power
7	Turn on the power switch. The device will start initial testing automatically
8	The interface will display after the initial testing is completed. 

Warning: for moving the instrument after use, refer to instructions in 4. Device Maintenance.

2. Device and Interface Display

2.1 Device and transfer cassette

eBlot™ L2 Fast Protein Transfer Device consists of one device and two transfer cassettes



Figure 1. eBlot L2 Device



Figure 2. eBlot L2 Back View




Figure 3. eBlot L2 Transfer Cassette, midi

2.2 Interface Display and Method Setting

2.2.1 Interface Display

eBlot™ L2 has three interfaces: main interface, method setting and system setting. The initial interface after powered on is the main interface which shows Channel A and B, and the selected methods.

Main Interface




Figure 4. Main Interface

Interface Notation	Description
AB	Channel A, Channel B
Select	Transfer method selection
Standard	Selected transfer method
MIDI / NONE	MIDI or NONE Transfer cassette
▶ / ◻	▶ Green indicates the channel is ready for run while gray indicates unavailable ◻ Transfer in progress or transfer has finished Press and hold it for 5 seconds to terminate the method if needed
Ready	Current channel status
⌂ 📄 ⚙	Interface navigation

Method Selection Interface

Press Select to enter the method selection interface.




Fig 5. Method selection interface for channel B

Interface notation	Description
Standard Short Long Method 4 Method 5 Method 6	List of method options, with the selected one highlighted
Buffer Cycles2Cycle Cycle15.0min Cycle25.0min Cooling Cycles0Cycle	Method settings
Channel B	Current channel
Select	Apply the selected method to the indicated channel and return to main interface
Edit	Edit the selected method. Method setting cannot be changed during transfer progress

Method Setting Interface

Navigate into  for method setting interface and select the method.

Tap  to edit the method settings.




Fig 6. Method Selection




Fig 7. Method Editing

Interface notation	Description
Buffer Cycles 2 Cycle 2	Choose number of transfer cycles
Cycle 1 5.0 min 5.0	Type in transfer time. Suggested <9 minutes for each transfer cycle
Cooling Cycles 0 Cycle 0	Choose number of cooling cycles
Cycle 0.0 min 0.0	Type in cooling time
Cooling buffer Water	Select cooling buffer type between transfer buffer and water(Distilled Water), If water is selected, fill the yellow cap container with water and connect it to the yellow outlet
	Rename the selected method
	Reset to default method setting
	Exit method editing and do not save the changes
	Save the method edits

System Setting Interface

Navigate into  for system settings.




Fig 8. System Setting

Notation	Description
Language 	Select between English/Chinese interface
Drain Residue Reagents 	Drain the residue reagents from the instruments tubing
Reset Pump Setting 	Reset to the default buffer inflow time
Reload Default Method 	Reset to the default method settings of Standard, Short, Long protocols
System Clean 	Perform system cleaning. Refer to the instructions in Section 4. Device Maintenance
 	Use by engineers only

2.2.2 Transfer Method Overview

eBlot™ L2 contains 3 default methods (Standard, Short, Long), and 3 customizable methods for various transfer needs.

eBlot L2 Default Methods

Standard

Standard	Buffer Cycles	2	Cycle
Short	Cycle1	5.0	min
Long	Cycle2	5.0	min
Method 4	Cooling Cycles	0	Cycle
Method 5			
Method 6			

Figure 9. Default Standard Method

Short

Standard	Buffer Cycles	1	Cycle
Short	Cycle1	5.0	min
Long	Cooling Cycles	0	Cycle
Method 4			
Method 5			
Method 6			

Figure 10. Default Short Method

Long

Standard	Buffer Cycles	3	Cycle
Short	Cycle1	5.0	min
Long	Cycle2	5.0	min
Method 4	Cycle3	5.0	min
Method 5	Cooling Cycles	0	Cycle
Method 6			

Figure 11. Default Long Method

3. Device Operation and Application Data

Please follow the instructions to obtain consistent and high quality data.

Crystallization may occur in reagents under low temperature. Please leave the reagents at room temperature to allow dissolution before use. Do not use expired or turbid reagents.

- Wear gloves to prevent contamination of gels, membranes and transfer sponges.
- Check the volume in the reagent bottle before each transfer to ensure there is adequate amount of transfer buffer (300 mL/Standard method).
- Use the provided waste container, tubing and caps. Regularly empty the waste container.
- Before transfer, cut off the gel well fingers and equilibrate the gel and membrane in the equilibrium buffer for 1 minute (60 mL/transfer).
- When assembling the transfer sandwich, place the gel and membrane close to the cassette bottom, with the large molecular weight proteins closer to the inlet/outlet.
- Gently remove bubbles between the gel and membrane using the roller dipped in the equilibrium buffer.
- Fully insert the transfer cassette into the channel with both hands until the scale lines of the cassette fully covered.
- After transfer, remove the transfer cassette from the instrument, and immediately place it in a tray containing distilled water to prevent the membrane from drying out.
- After each transfer, rinse the cassette thoroughly under running water to cool down the cassette and to prevent any salt precipitation.

3.1 Reagent Preparation

Before you use the proprietary transfer reagent kit (Cat. No. L01015-30) on eBlot L2, dilute the concentrated transfer buffer (Cat. No. B0056) and the concentrated equilibrium buffer (Cat. No. B0057) according to the protocol below.

Ingredients for 1X transfer buffer	Volume
eBlot L2 Concentrated Transfer Buffer, 5X	1 L
ddH ₂ O	4 L
Total	5 L
Mix well before use	

Ensure the transfer buffer is adequate before each transfer and do not reuse (300 mL/Standard method).

Ingredients for 1X equilibrium buffer	Volume
eBlot L2 Concentrated Equilibrium Buffer, 5X	100 mL
ddH ₂ O	400 mL
Total	500 mL
Mix well before use	

Do not reuse the equilibrium buffer (60 mL/transfer).

3.2 Method Selection Guide

3.2.1 Turn on the device by pressing the power switch on the back of the device.

3.2.2 The main interface will display Channel A and B, and the current methods.

3.2.3 Select the transfer method for your target protein size.

Refer to 2.2 Interface Display and Method Setting for operation details.

Transfer Method Selection Guide

Default Transfer Method	Protein Size Range
Short	< 20 kDa
Standard	20-200 kDa
Long	> 200 kDa

Note:

- 0.22 µm PVDF/NC membranes are recommended for proteins < 20 kDa.
- If you are using GenScript SurePAGE Mini or Midi Gel 4-12% or 4-20%, please select Standard method.
- Users can customize the transfer methods to optimize the transfer efficiency.
 - If the gel concentration and the protein size are lower, transfer time and cycles can be reduced.
 - If the gel concentration and the protein size are higher, transfer time and cycles can be increased.

3.3 Operation Instructions

1. Prepare 60 mL 1X equilibrium buffer in the provided plastic container.

2. Prepare 250 mL distilled water in the provided tray.

3. Ensure adequate transfer buffer and regularly empty the waste container.

Note: Each Standard method consumes approximately 300 mL 1X transfer buffer.

4. Place the transfer cassette on the bench and open the cassette as shown in Figure 12.

a. Release the cassette latch by rotating it upward.

b. Lift the latch and fully open the cassette cover to approximately $>90^\circ$.

c. Note the anode side of the transfer cassette is marked with a red sealing strip.




Figure 12. Opening the transfer cassette

5. Cut off the gel well fingers as shown in Figure 13.

Note:

a. Rinse the gel cassette with water to remove any remaining running buffer.

b. Cut off any uneven part of the gel, and the stacking gel of the self-casted gel.




Figure 13. Cut off gel well fingers

6. Place the membrane and gel with wells cut off in the equilibrium buffer for 1 minute*.

Note: Activate the PVDF membrane in applicable alcohol before equilibrium.

*The gel equilibrium time should not exceed 5 minutes to prevent protein dispersion.

7. Place one dry sponge on the red anode side.

Note:

- a. The anode side is marked with red sealing strip.
- b. Use dry sponges for transfer. Sponges should NOT be wet.
- c. Sponges are NOT reusable and should be discarded after the transfer.




Figure 14. Placing transfer sponge

8. Take out the membrane from equilibrium buffer and place on top of the dry sponge, then place the gel on top of the membrane with large proteins towards the inlet/outlet.

Note:

- a. Correct size of the components: $\text{Sponge} \geq \text{Membrane} \geq \text{Gel}$.
- b. Place gel and membrane towards the bottom, and do not exceed the edge of sponge as shown in Figure 15.
- c. Avoid overlapping or contact between two mini gels.
- d. Finish the assembly of transfer sandwich quickly to prevent the gel and membrane from drying out.
- e. Do not place gel and membrane above the green dotted line.
- f. When transfer only 1 mini gel in the midi cassette, place the mini gel in the middle of the same indicated area.



a. Placing 2 mini gels

b. Placing 1 midi gel

Figure 15. Placing gel and membrane

9. Gently remove bubbles between the gel and membrane using the roller dipped in the equilibrium buffer.

Note: Do not press the gel to cause gel deformation.

10. Place another dry sponge on top of gel.

Note: Do not frequently remove the top dry sponge to create bubbles between gel and membrane.




Figure 16. Assembly of transfer sandwich

11. Close the transfer cassette as shown in Figure 17.

a. Close the top cover of transfer cassette and press firmly.

b. Rotate the latch upward to lock the cassette.

Figure 17c shows the fully closed transfer cassette




Figure 17. Closing the transfer cassette

12. Fully insert the assembled transfer cassette into the channel until the scale lines of the cassette completely covered as shown in Figure 18.

Note:

- a. Note the proper orientation and loading of the cassette. The scale lines of the cassette should be facing towards the user.
- b. Start transfer within 5 minutes after the assembly of transfer sandwich.




Figure 18. Inserting the transfer cassette

13. Press  for the corresponding channel to initiate the transfer method. The screen will display a countdown timer as shown in Figure 19.




Figure 19. Start the transfer

14. The device will beep when the countdown reaches 00:00 as shown in Figure 20. Press  the stop button to return to the initial interface.




Figure 20. Transfer completed

15. Remove the transfer cassette from the channel.

Note: Slightly heated transfer cassette is normal.

16. Open the cassette according to Step 4. Immediately place the transfer sandwich into a tray containing distilled water. Disassemble the sandwich and take out the membrane for the next steps.

Note: If PVDF membrane is dry after transfer, it must be reactivated in applicable alcohol for 1 minute before proceeding to the subsequent experiments.





Figure 21. Disassembly of the transfer sandwich

After each transfer, rinse the cassette with running water for 1 minute and set dry on the bench.

3.4 Application Data

Side-by-side comparison data between eBlot™ L2 Standard Method and traditional wet transfer using GenScript 4-12% Precast Gel.




Target protein	Samples	Primary Ab	Secondary Ab
EGFR	HEK293 Cell Lysate, left to right : 20ug, 10ug, 5ug, 2.5ug, 1.25ug, 0.62ug, 0.31ug, 0.15ug	Recombinant Anti-EGFR antibody, mAb, rabbit (Abcam, ab52894), 1:1000	Anti-rabbit IgG (HRP) (GOAT) antibody peroxidase (Rockland, 610-103-122), 1:10000
β-actin	HEK293 Cell Lysate, left to right : 40ug, 20ug, 10ug, 5ug, 2.5ug, 1.25ug, 0.62ug, 0.31ug, 0.15ug	THE™ beta Actin antibody, mAb, mouse (Genscript, A00702), 1:5000	Anti-mouse IgG (HRP) (GOAT) antibody peroxidase (Rockland, 610-103-121), 1:10000
GAPDH	HEK293 Cell Lysate, left to right : 40ug, 20ug, 10ug, 5ug, 2.5ug, 1.25ug, 0.62ug, 0.31ug, 0.15ug	GAPDH antibody, mAb, mouse (GenScript, customized), 1:1000	Anti-mouse IgG (HRP) (GOAT) antibody peroxidase (Rockland, 610-103-121), 1:10000
Histone H3	HEK293 Cell Lysate, left to right : 40ug, 20ug, 10ug, 5ug, 2.5ug, 1.25ug, 0.62ug, 0.31ug, 0.15ug	Histone H3 antibody, pAb, rabbit (GenScript, customized), 1:1000	Anti-rabbit IgG (HRP) (GOAT) antibody peroxidase (Rockland, 610-103-122), 1:10000



4. Device Maintenance

Regular maintenance is highly recommended to maintain the upkeep of the device and performance of protein transfer.

Components	Maintenance guide										
Transfer cassette	After each transfer, rinse the cassette with water for 1 minute and set dry on the bench.										
Transfer channels and Tubing	Perform system cleaning once every 100 transfers, or once a month for higher usage frequency.										
	1. Prepare 80 mL cleaning buffer to clean both channels according to the protocol below.										
	<table><tr><th>Ingredients</th><th>Volume</th></tr><tr><td>Ethanol (>99%)</td><td>24 mL</td></tr><tr><td>Distilled water</td><td>56 mL</td></tr><tr><td>Total</td><td>80 mL</td></tr><tr><td colspan="2">Mix well before use</td></tr></table>	Ingredients	Volume	Ethanol (>99%)	24 mL	Distilled water	56 mL	Total	80 mL	Mix well before use	
	Ingredients	Volume									
	Ethanol (>99%)	24 mL									
Distilled water	56 mL										
Total	80 mL										
Mix well before use											
Note: Excessive volume may result in residual cleaning solution.											
2. Insert empty cassettes into both channels. Drain residue reagents from tubing by selecting the "Drain" function on the "Setting" interface.											
3. Disconnect only "Waste 2" tubing from the waste container cap and insert it into a container filled with cleaning solution. Keep "Waste 1" on the waste container cap.											
4. Connect the yellow tubing to the distilled water inlet, and insert the other end of the tubing to the container with distilled water.											
5. Press System Clean button to run the cleaning program for 3 minutes.											
6. After cleaning is completed, reconnect the "Waste 2" tubing to the waste container.											
Drain the tubing	Press Drain to flush the tubing.										
Others	Device left unused for long-term										
	If the device will not be used for an extended period (>2 weeks), flush the tubing twice, turn off the device, and seal the transfer solution bottle with a closed cap to prevent liquid evaporation.										
	Moving the device										
	When moving the device, please ensure that the main interface is functional when powered on, and no liquid in the channels. Keep the instrument levelled during moving to prevent liquid from spilling out of the channels or tubing.										

5. Troubleshooting

Common Issues	Causes	Resolution
Q1 Strong transfer background or contamination	1. Contamination or bacterial growth in the gel running buffer. 2. Contamination in the PVDF membrane activation reagent (methanol, ethanol, isopropanol). 3. Contamination in the eBlot L2 equilibrium buffer. 4. Contamination due to the reuse of transfer sponges. 5. Residue on the eBlot L2 transfer cassette. 6. Lack of instrument cleaning.	1. Rinse and dry the electrophoresis tank after use; do not reuse gel running buffer. 2. Do not reuse PVDF membrane activation reagent. 3. Do not reuse eBlot L2 equilibrium buffer. 4. Do not reuse transfer sponges. 5. Rinse the transfer cassette for 1 minute and set dry after each transfer. 6. Refer to Section 4. Device Maintenance to regularly perform system cleaning.
Q2 Irregular bands along the membrane edge	1. Two mini gels are overlapping or in contact. 2. Gel exceeds the edge of the membrane in the transfer sandwich. 3. Gel or membrane exceeds the edge of the transfer surface of the cassette.	1. Avoid overlapping and contact between 2 mini gels. 2. The correct size of transfer sandwich components: Sponge > Membrane > Gel. 3. Place gel and membrane within the indicated area.  Transfer 2 mini gels in one cassette  Transfer 1 midi gel in one cassette
Q3 Lost bands on PVDF membrane after transfer and Ponceau staining	1. PVDF membrane is not activated before transfer. 2. PVDF membrane is dry and deactivated after transfer.	PVDF membrane needs to be activated in the applicable alcohol before transfer, and reactivated after transfer if the PVDF membrane is dry.  Membrane Inactivation Ponceau Staining  Reactivation Followed by Ponceau Staining

Common Issues	Causes	Resolution
Q4 Low signal for large protein bands	1. Inappropriate membrane pore size 2. Transfer time is too short 3. Gel concentration is not appropriate	1. Use 0.45 μm PVDF/NC membrane if signal is low for proteins >150 kDa 2. Increase transfer time or cycles 3. Choose lower concentration gels
		
Q5 Low signal for small protein bands	1. Inappropriate membrane pore size 2. Transfer time is too long 3. Protein dispersion 4. Gel concentration is not appropriate	1. Use 0.22 μm PVDF/NC membrane if signal is low for proteins <20 kDa 2. Reduce transfer time or cycles 3. Start transfer immediately after electrophoresis; reduce the gel equilibrium time 4. Choose higher concentration gels
		
Q6 Lost bands for target proteins	1. The assembly of transfer sandwich takes too long that causes dryness or deactivation of the membrane 2. Did not start transfer immediately after the assembly of transfer cassette 3. Degradation of target proteins 4. Same as Q3, Q4 or Q5	1. Complete the assembly of transfer cassette quickly; re-equilibrate or reactivate the membrane if needed. 2. Start the transfer within 5 minutes after the assembly of transfer sandwich 3. Prepare new protein samples 4. Same as Q3, Q4 or Q5
Q7  Current low: Channel A. Please check buffer dilution. 	1. Incorrect transfer buffer dilution: transfer buffer concentration is too low 2. Incorrect assembly of transfer sandwich 3. PVDF membrane is not activated	1. Prepare and dilute the transfer buffer to 1X according to the protocol in 3.1 Reagent Preparation 2. Correctly assemble the transfer sandwich according to 3.3 Operation Instructions 3. Activate PVDF membrane in applicable alcohol before transfer

Common Issues	Causes	Resolution
Q8  Current over limit: Channel A. Please check buffer dilution. Ok	1. The surface temperature of transfer cassette is too high due to repetitive transfers or insufficient rinse between two transfers 2. Incorrect transfer buffer dilution: transfer buffer concentration is not diluted or too high 3. Abnormal substances exist inside the transfer cassette that causes short circuits between anode and cathode 4. Incorrect assembly of transfer sandwich 5. Waste container is too full that causes reverse flow during transfer	1. After each transfer, rinse the transfer cassette in water for 1 minute and cool it down to room temperature 2. Prepare and dilute the transfer buffer to 1X according to the protocol in 3.1 Reagent Preparation 3. Inspect if there are abnormal substances inside the transfer cassette; contact GenScript Tech Support if not resolved 4. Properly assemble the transfer sandwich 5. Regularly empty the waste container
Q9  Buffer low: Channel B. Check the liquid level in the buffer bottle. Ok	1. Insufficient transfer buffer 2. Incorrect connection to the transfer buffer container, or twisted/broken tubing 3. Crystallization of reagent clogs in the tubing 4. Tubing end is not fully submersed in the buffer 5. Tubing is disconnected from the buffer bottle cap 6. Same as Q7	1. Prepare adequate amount of transfer buffer (300ml/standard transfer) before each transfer 2. Check the transfer buffer tubing 3. If the device is left unused for an extended period, empty the tubing to prevent crystallization; if the tubing is clogged, perform system cleaning (4.0 Device Maintenance) 4. Make sure the end of tubing is submerged in the buffer and the tubing connects are intact 5. Check the tubing connection under the bottle cap 6. Same as Q7
Q10  Removal of cassette detected Ok	Channel A transfer cassette is accidentally removed, and program is terminated	The transfer program is automatically terminated. Insert the transfer cassette and restart the method Note: Removing the transfer cassette during the liquid inflow process may result in a small amount of transfer buffer spraying out, triggering a liquid leakage warning

Common Issues	Causes	Resolution
Q11  Leakage detected	1. The transfer cassette is not fully inserted into the instrument 2. The sealing ring at the connection between the transfer cassette and the instrument is aged, causing liquid leakage 3. Removing the transfer cassette during the liquid inflow process may result in a small amount of transfer buffer spraying out, triggering a liquid leakage warning	Turn off the instrument and disconnect the power. Remove the black soft plug from the bottom of the instrument and elevate the rear of the instrument to allow the liquid inside the instrument to drain out Note: Do not invert the instrument. Troubleshoot according to the causes. If the liquid leakage error persists after restarting the instrument, please contact GenScript Technical Support
Q12  Liquid sensor error		
Q12  Pump error		
	Hardware malfunction	Please contact GenScript Tech Support
Q12  Valve error		
Q12  Integrated driver error		

6. Consumables Offering


eBlot L2 Transfer Reagent Kit

Cat. No.	Product Name	Size	Description
L01015-30	eBlot L2 Transfer Basic Kit, Midi	1 Kit	eBlot L2 concentrated transfer reagents for 30 midi or 60 mini transfers with standard method

L01015-30, eBlot L2 Transfer Basic Kit, Midi

Component name	Size	Quantity
eBlot L2 Concentrated Transfer Buffer, 5X	1.8 L	1
eBlot L2 Equilibration Buffer (5X)	0.4 L	1
eBlot L2 Transfer Sponge, Midi	15 pk	2

7. Related Products



SurePAGE™ Bis-Tris Gel, Midi, 10pk/box

Cat. No.	Concentration	Well format
M00995	4-12%	20
M00996	4-12%	26

SurePAGE™ Bis-Tris Gel, Mini, 10pk/box

Cat. No.	Concentration	Well format
M00652	4-12%	10
M00653	4-12%	12
M00654	4-12%	15
M00655	4-20%	10
M00656	4-20%	12
M00657	4-20%	15
M00658	8-16%	10
M00659	8-16%	12
M00660	8-16%	15
M00661	8%	10
M00662	8%	12
M00663	8%	15
M00664	10%	10
M00665	10%	12
M00666	10%	15
M00667	12%	10
M00668	12%	12
M00669	12%	15

Other Electrophoresis Products

Cat. No.	Product Name	Size
L01001	GenBox Midi Electrophoresis Tank	1 unit
L01021	GenBox Mini Plus Electrophoresis Tank	1 unit
L01022	GenBox Midi/Mini Electrophoresis Tank	1 unit
L00780	GenBox Mini Electrophoresis tank	1 unit
L00781	GenBox Mini Blot module	1 unit
L00782	GenBox Mini Blot System	1 unit
M00138	Tris-MOPS-SDS Running Buffer Powder	5 PK/Box
M00677	MES SDS Running Buffer Powder	5 PK/Box
M00624	Broad Multi Color Pre-Stained Protein Standard	250 µl, 1250 µl
M00516	PAGE-MASTER Protein Standard (for SDS-PAGE)	500 µl
M00521	WB-MASTER Protein Standard	250 µl
M00676	4X LDS Sample Buffer	10 ml, 250 ml
MB01015	5X Sample Buffer	5 ml
M00139	Transfer Buffer Powder	10 PK/Box

Western Blotting Products

Cat. No.	Product Name	Size
L00657	eStain™ L1	1 unit
L00753	eStain L1C Protein Staining Kit	1 unit
L00816	eZwest™ Lite Automated Western Device	1 unit
L00818	eZwest Diluent Kit	1 unit
A00702	THE™ beta Actin Antibody, mAb, Mouse	100 µg

8. Warranty

GenScript warrants that eBlot™ L2 to be free from defects in materials and workmanship under normal use and service for a period of twelve (12) months from the date of installation by GenScript or its authorized distributor.

GenScript agrees within the warranty period, at its discretion, to repair defects in the material or workmanship or to furnish a repaired or refurbished product of equal value in exchange without a fee. Such repairs require verification of the defect or malfunction and proof of purchase as confirmed by showing the serial number with the original dated sales receipt.

Warranty Limitations:

The warranty excludes failure resulting from the following situations:

- Instrument failure caused by improper handling or improper human operation
- Repairs or modifications by any party other than GenScript or an authorized agent
- Use of accessories or other spare parts supplied by any party other than GenScript
- Damage caused by natural disasters (e.g. earthquakes, landslides, hurricanes, tornadoes, etc.)
- Damage caused by the use of improper reagents (except for chemicals recommended by the instrument manufacturer)

For consulting or repair services, please contact Genscript customer service after confirming the instrument model, instrument serial number, order number, purchase date and other relevant information.

- Instrument model: _____
- Instrument serial number: _____
- Order number: _____
- Date of purchase: _____

USA

Tel: +1-732-885-9188

Toll-Free: +1-877-436-7274

Fax: +1-732-210-0262

Email: support.US@genscript.com

Address: 860 Centennial Ave. Piscataway NJ 08854

China

Tel: +86-025-58897288 ext 5810

Fax: +86-025-58897288-5815

Email: product@genscript.com.cn

Address: 28 YongXi Road, JiangNing District, Nanjing, China

Asia/Pacific

Tel: +65 3159 1898

Email: sales.AP@genscript.com

Fax: +65 6491 5073

Address: 164 Kallang Way, #06-12, Singapore 349248

Europe

Tel: +31 (0) 71 569 0120

Email: support.EU@genscript.com

Address: Treubstraat 1, 2288Eg Rijswijk Zh Netherlands